

 CELLSTORE	Document No.: QM12-013-02-A04
Document Name: Certificate of Analysis	

Certificate of Analysis

Report No.: SC-N-2026001

Product	Cryopreservation Medium	Product Code	CS-SC-N1
Product Specification	100mL/Bottle	Lot Number	01251211
Manufacture Date	2025.12.17	Expiration Date	2027.06.16
Storage Condition	2-8°C, protect from light		
Basis of Testing	Quality Specification for Stem Cell Cryopreservation Medium		
Test	Acceptance Criteria		Results
Appearance	Colorless, pale yellow or yellow transparent liquid		Colorless
pH	7.2±0.5		7.0
Osmolality	1460~1560mOsmol/kg		1537mOsmol/kg
Particulate Matter	Each container should meet the requirements: ≥10 μm particle: ≤ 25 particles/mL ≥25 μm particle: ≤ 3 particles/mL		≥10 μm particle: 1 particles/mL ≥25 μm particle: 1 particles/mL
Endotoxin	<0.125 EU/mL		Conforms
Sterility	Sterile		Conforms
Cell Viability	≥85%		90%
Conclusion	☒ Qualified ☐ Unqualified		
Annotation	N/A		

Reported by: Liang Hui

Date: 2026.01.06

Reviewed by: Wu Yingge

Date: 2026.01.06

Approved by: Gao Haifeng

Date: 2026.01.06

